

Blood Creatinine Estimation

Experiment 9 - Practical Notes

Name: _____ Roll no.: _____ Date: _____

Aim

To determine creatinine by the prescribed creatinine assay.

Principle

A validated creatinine method relates a corrected colour or enzyme response to a known standard. Interferences and measurement timing depend on the selected method.

Requirements

- Assigned test material and the prescribed method reagents or standards.
- Calibrated volumetric glassware, appropriate instrument and record sheet.

Procedure

1. Prepare and identify the test

Confirm the analyte, method version, concentration basis and test-portion amount. Prepare the test solution as specified and record any dilution.

2. Set the measurement basis

Check the instrument or standardised titrant. Prepare the specified blank, standard and control so that the test and comparison use the same basis.

3. Make the measurement

Apply the prescribed the prescribed creatinine assay method and record test, standard and control responses. Record replicate readings and any interfering appearance.

4. Calculate

Apply the stated equation with the correct blank correction, dilution and unit conversion. Show the substitution and retain the observed readings.

5. Assess the result

Compare replicate agreement and the assigned specification. Record a result with units and the limitations of the method.

Observations and Record

Measurement	Reading
Test portion / dilution	_____
Blank / standard	_____
test, standard and control responses	_____
Calculated content and unit	_____

Worked Example

The numbers below are illustrative; enter your own observations in the record.

- $\text{Concentration} = \text{corrected test response} \div \text{corrected standard response} \times \text{standard concentration} \times \text{dilution factor}$.
- A ratio of 0.65 against a 2 mg/dL standard gives 1.30 mg/dL in an undiluted teaching calculation.

Result

Record the completed observations, calculations and conclusion with units. State any limitation of the observation.

Precautions

- Use the exact method reagents and assigned endpoint; different methods are not interchangeable.
- Record the blank correction and the actual test portion rather than an assumed value.
- Hazardous reagent operations require the specified laboratory facilities and supervision.

Viva Questions and Answers

1. What produces the analytical signal?

A validated creatinine method relates a corrected colour or enzyme response to a known standard. Interferences and measurement timing depend on the selected method.

2. Why include a blank or control?

It identifies contributions unrelated to the analyte or checks that the measurement system performed appropriately.

3. What must accompany the result?

Units, method, dilution basis, relevant control performance and any limitation.

Continue Learning with FirstHope

Complete Notes on the FirstHope App

Complete subject and semesters are available on FirstHope app only.

- Get complete subject and semester notes in the FirstHope app.
- Available for B.Pharm, D.Pharm, M.Pharm and Pharm.D students.

[Click here to get the FirstHope app](#)

Join the FirstHope Telegram Channel

Join FirstHope Notes and keep up with channel updates.

[Click here to join Telegram](#)

Instagram

@firsthope.official

[Click here to follow Instagram](#)

YouTube Channel

FirstHope-Pharmacy / Live streams

[Click here to watch the YouTube channel](#)

WhatsApp Channel

Follow FirstHope on WhatsApp.

[Click here to join the WhatsApp channel](#)

FirstHope Website

firsthopenotes.com

[Click here to visit the website](#)